



## Certificate of Calibration

41 Waukegan Rd  
Lake Bluff, IL 60044  
Ph. 847.295.4542  
Fax. 847.295.7954

Customer Site

ECHO INC.  
400 OAKWOOD RD.  
  
LAKE ZURICH IL 60047



Certificate - 2578-1

*Original*

This certifies that the Hardness Tester listed has been calibrated in accordance with **ASTM E384-11e1** verification procedures and incorporating the requirements of ISO 17025-2005. Stage Micrometers used for measuring eyepiece are traceable to N.I.S.T. and all test blocks used were standardized on a verified tester with loads traceable to N.I.S.T. Uncertainty was calculated using a factor of K=2 with a confidence level of 95%.

Model Number	TUKON 1202	Location Calibrated	Customer Site
Serial Number	1202-02-0170	Condition of Equipment as Found	Good
Asset #/PO #	PH86914 N/A	Indenter(s) Vickers	EV002873 Knoop

### Measuring Microscope

These measurements are not part of the ASTM E384-11e1 indirect method and are for reference only. Therefore, this section of the report is not accredited by A2LA.

Stage Micrometer #	14504	with a	50X	Objective			
Nominal Length ( $\mu\text{m}$ )	0	20	60	100	200	Result	
Readings as found ( $\mu\text{m}$ )	0.00	20.00	59.60	99.70	199.10	Pass	
Readings as left ( $\mu\text{m}$ )	0.00	20.00	59.60	99.70	199.10	Pass	
$\pm$ Tolerance ( $\mu\text{m}$ )	0.4	0.4	0.4	0.5	1		
Uncertainty	N/A	N/A	N/A	N/A	N/A		

### Environmental Data

Start		Finish
Temp (F.) 71	Humidity 21.7%	Temp (F.) 71 Humidity 26.6%

### Summary of Hardness Tester Results

 See page 5 for additional notes and information on any deviations.

Test Block	Readings as Found				Readings as Left			
	Tol +/-	Actual	Error %	Pass/Fail	Actual	Error %	Pass/Fail	Uncertainty
718HV0.3	31	742	1.51	Pass	742	1.51	Pass	27
500HV1.0	20	495	0.54	Pass	495	0.54	Pass	14
-	-	-	-		-	-		-
-	-	-	-		-	-		-

X

Report Date: January 29, 2016  
Requested Due Date: January 28, 2017


**Data Sheet  
(As Found)**

**Certificate - 2578-1**

Stage Micrometer #	14504	with a	50X	Objective			
Nominal Length (μm)	0	20	60	100	200	Result	
Readings as Found (μm)	0.00	20.00	59.60	99.70	199.10	Pass	
± Tolerance (μm)	0.4	0.4	0.4	0.5	1.0		
Result	Pass	Pass	Pass	Pass	Pass		


Test Block # 1	39LS
Ave Hardness	718.0
Load (gf)	300
Indenter type	Vickers
Diagonals	27.8
Uncertainty (a)	0.22
Resolution (b)	0.10

	D <sub>2</sub>	D <sub>1</sub>	Average	Hardness
1	27.5	27.5	27.50	735
2	27.5	27.4	27.45	738
3	27.2	27.7	27.45	738
4	26.6	27.4	27.00	763
5	27.6	27.4	27.50	735
	Average		27.38	742
	Error (μ)		0.42	Pass
	Repeatability		1.83%	Pass
	Result		Pass	
	Range (c)		0.5	

U <sub>1</sub>	U <sub>2</sub>	U <sub>3</sub>	Combined Uncertainty	Expanded Uncertainty (mm)	Expanded Uncertainty (%)
	$\frac{b}{1.732} (1000)$		$\sqrt{(U_1)^2 + (U_2)^2 + (U_3)^2}$	Combined Uncertainty * 2	
0.00011	5.77367E-05	0.000500	0.000515202	0.001030405	3.71%
					Vickers 27

Test Block # 2	63LR
Ave Hardness	500.0
Load (gf)	1000
Indenter type	Vickers
Diagonals	60.9
Uncertainty (a)	0.44
Resolution (b)	0.10

	D <sub>2</sub>	D <sub>1</sub>	Average	Hardness
1	61.0	60.8	60.90	500
2	61.8	61.6	61.70	487
3	61.2	61.1	61.15	496
4	61.2	61.1	61.15	496
5	61.2	61.3	61.25	494
	Average		61.23	495
	Error (μ)		0.33	Pass
	Repeatability		1.31%	Pass
	Result		Pass	
	Range (c)		0.8	

U <sub>1</sub>	U <sub>2</sub>	U <sub>3</sub>	Combined Uncertainty	Expanded Uncertainty (mm)	Expanded Uncertainty (%)
	$\frac{b}{1.732} (1000)$		$\sqrt{(U_1)^2 + (U_2)^2 + (U_3)^2}$	Combined Uncertainty * 2	
0.00022	5.77367E-05	0.000800	0.000831705	0.00166341	2.73%
					Vickers 14


The reported uncertainty uses a coverage factor of K=2, representing a confidence level of 95%.

**Data Sheet  
(As Found)**

**Certificate - 2578-1**


Test Block # 3	
Ave Hardness	
Load (gf)	
Indenter type	
Diagonals	
Uncertainty (a)	
Resolution (b)	

	D <sub>2</sub>	D <sub>1</sub>	Average	Hardness
1				
2				
3				
4				
5				
		Average		
		Error (μ)		
		Repeatability		
		Result		
		Range (c)		

U <sub>1</sub>	U <sub>2</sub>	U <sub>3</sub>	Combined Uncertainty	Expanded Uncertainty (mm)	Expanded Uncertainty (%)
	$\frac{b}{1.732} (1000)$		$\sqrt{(U_1)^2 + (U_2)^2 + (U_3)^2}$	$Combined\ Uncertainty * 2$	
-	-	-	-	-	-
				0	#VALUE!

Test Block # 4	
Ave Hardness	
Load (gf)	
Indenter type	
Diagonals	
Uncertainty (a)	
Resolution (b)	

	D <sub>2</sub>	D <sub>1</sub>	Average	Hardness
1				
2				
3				
4				
5				
		Average		
		Error (μ)		
		Repeatability		
		Result		
		Range (c)		

U <sub>1</sub>	U <sub>2</sub>	U <sub>3</sub>	Combined Uncertainty	Expanded Uncertainty (mm)	Expanded Uncertainty (%)
	$\frac{b}{1.732} (1000)$		$\sqrt{(U_1)^2 + (U_2)^2 + (U_3)^2}$	$Combined\ Uncertainty * 2$	
-	-	-	-	-	-
				0	#VALUE!

The reported uncertainty uses a coverage factor of K=2, representing a confidence level of 95%.

**Data Sheet  
(As Left)**

**Certificate - 2578-1**

Stage Micrometer #	14504	with a	50X	Objective			
<b>Nominal Length (μm)</b>	0	20	60	100	200	<i>Result</i>	
<b>Readings as Left (μm)</b>	0.00	20.00	59.60	99.70	199.10	<b>Pass</b>	
± Tolerance (μm)	0.4	0.4	0.4	0.5	1.0		
<b>Result</b>	Pass	Pass	Pass	Pass	Pass		

Test Block # 1	39LS
Ave Hardness	718.0
Load (gf)	300
Indenter type	Vickers
Diagonals	27.8
Uncertainty (a)	0.22
Resolution (b)	0.10

	D <sub>2</sub>	D <sub>1</sub>	Average	Hardness
1	27.5	27.5	27.50	735
2	27.5	27.4	27.45	738
3	27.2	27.7	27.45	738
4	26.6	27.4	27.00	763
5	27.6	27.4	27.50	735
	Average		27.38	742
	Error (μ)		0.42	Pass
	Repeatability		1.83%	Pass
	Result		Pass	
	Range (c)		0.5	

U <sub>1</sub>	U <sub>2</sub>	U <sub>3</sub>	Combined Uncertainty	Expanded Uncertainty	Expanded Uncertainty
	$\frac{b}{1.732} (1000)$		$\sqrt{(U_1)^2 + (U_2)^2 + (U_3)^2}$	(mm)	(%)
0.00011	5.77367E-05	0.0005	0.000515202	0.001030405	3.71%
					Vickers 27

Test Block # 2	63LR
Ave Hardness	500
Load (gf)	1000
Indenter type	Vickers
Diagonals	60.9
Uncertainty (a)	0.44
Resolution (b)	0.10

	D <sub>2</sub>	D <sub>1</sub>	Average	Hardness
1	61.00	60.80	60.90	500
2	61.80	61.60	61.70	487
3	61.20	61.10	61.15	496
4	61.20	61.10	61.15	496
5	61.20	61.30	61.25	494
	Average		61.23	495
	Error (μ)		0.33	Pass
	Repeatability		1.31%	Pass
	Result		Pass	
	Range (c)		0.80	

U <sub>1</sub>	U <sub>2</sub>	U <sub>3</sub>	Combined Uncertainty	Expanded Uncertainty	Expanded Uncertainty
	$\frac{b}{1.732} (1000)$		$\sqrt{(U_1)^2 + (U_2)^2 + (U_3)^2}$	(mm)	(%)
0.00022	5.77367E-05	0.0008	0.000831705	0.00166341	2.73%
					Vickers 14


The reported uncertainty uses a coverage factor of K=2, representing a confidence level of 95%.

**Data Sheet  
(As Left)**

**Certificate - 2578-1**


Test Block # 3	
Ave Hardness	
Load (gf)	
Indenter type	
Diagonals	
Uncertainty (a)	
Resolution (b)	

	D <sub>2</sub>	D <sub>1</sub>	Average	Hardness
1				
2				
3				
4				
5				
		Average		
		Error (μ)		
		Repeatability		
		Result		
		Range (c)		

U <sub>1</sub>	U <sub>2</sub>	U <sub>3</sub>	Combined Uncertainty	Expanded Uncertainty (mm)	Expanded Uncertainty (%)
	$\frac{b}{1.732} (1000)$		$\sqrt{(U_1)^2 + (U_2)^2 + (U_3)^2}$	Combined Uncertainty * 2	
-	0.000058	-	-	-	-
				0	#VALUE!

Test Block # 4	
Ave Hardness	
Load (gf)	
Indenter type	
Diagonals	
Uncertainty (a)	
Resolution (b)	

	D <sub>2</sub>	D <sub>1</sub>	Average	Hardness
1				
2				
3				
4				
5				
		Average		
		Error (μ)		
		Repeatability		
		Result		
		Range (c)		

U <sub>1</sub>	U <sub>2</sub>	U <sub>3</sub>	Combined Uncertainty	Expanded Uncertainty (mm)	Expanded Uncertainty (%)
	$\frac{b}{1.732} (1000)$		$\sqrt{(U_1)^2 + (U_2)^2 + (U_3)^2}$	Combined Uncertainty * 2	
-	-	-	-	-	-
				0	#VALUE!

The reported uncertainty uses a coverage factor of K=2, representing a confidence level of 95%.

**Additional Comments:**